

Federal Communications Commission

§ 18.207

Charge when additional time is required to put into effect the corrective measures or to complete the investigation. The request for extension of time shall be accompanied by a progress report showing what has been accomplished to date.

§ 18.119 Importation.

ISM equipment shall be refused entry or withdrawal for consumption into the Customs territory of the United States, unless accompanied by a copy of FCC Form 740, in accordance with the provisions of subpart K, part 2 of this chapter.

§ 18.121 Exemptions.

Non-consumer ultrasonic equipment, and non-consumer magnetic resonance equipment, that is used for medical diagnostic and monitoring applications is subject only to the provisions of §§ 18.105, 18.109 through 18.119, 18.301 and 18.303 of this part.

[59 FR 39472, Aug. 3, 1994; 60 FR 47302, Sept. 12, 1995]

Subpart B—Applications and Authorizations

§ 18.201 Scope.

This subpart contains the procedures and requirements for authorization to market or operate ISM equipment under this part.

§ 18.203 Equipment authorization.

(a) Consumer ISM equipment, unless otherwise specified, must be authorized under either the Declaration of Conformity or certification procedure prior to use or marketing. An application for certification shall be filed with the Commission on an FCC Form 731, pursuant to the relevant sections in part 2, subpart J of this chapter and shall also be accompanied by:

(1) A description of measurement facilities pursuant to § 2.948, or reference to such information already on file with the Commission.

(2) A technical report pursuant to §§ 18.207 and 18.311.

(b) Consumer ultrasonic equipment generating less than 500 watts and operating below 90 kHz, and non-consumer ISM equipment shall be subject

to verification, in accordance with the relevant sections of part 2, subpart J of this chapter.

(c) Grants of equipment authorization issued, as well as on-site certifications performed, before March 1, 1986, remain in effect and no further action is required.

[50 FR 36067, Sept. 5, 1985, as amended at 63 FR 36603 July 7, 1998]

§ 18.207 Technical report.

When required by the Commission a technical report shall include at least the following information:

(a) A description of the measurement facilities in accordance with § 2.948. If such a description is already on file with the Commission, it may be included by reference.

(b) A copy of the installation and operating instructions furnished to the user. A draft copy of such instructions may be submitted with the application, provided a copy of the actual document to be furnished to the user is submitted as soon as it is available, but no later than 60 days after the grant of the application.

(c) The full name and mailing address of the manufacturer of the device and/or applicant filing for the equipment authorization.

(d) The FCC Identifier, trade name(s), and/or model number(s) under which the equipment is or will be marketed.

(e) A statement of the rated technical parameters that includes:

(1) A block and schematic diagram of the circuitry.

(2) Nominal operating frequency.

(3) Maximum RF energy generated.

(4) Electrical power requirements of equipment.

(5) Any other pertinent operating characteristics.

(f) A report of measurements, including a list of the measuring equipment used, and a statement of the date when the measuring equipment was last calibrated and when the measurements were made. The frequency range that was investigated in obtaining the report of measurements shall be indicated. See also §§ 18.309 and 18.311.

[50 FR 36067, Sept. 5, 1985, as amended at 63 FR 36603, July 7, 1998]